

SAMPLE

検査成績表 TEST REPORT

製品名称
PRODUCT NAME

デジタルマルチメータ
DIGITAL MULTIMETER

形名
MODEL TY720

計器番号
SERIAL No.

手配 No.
ORDER No.

タグ No.
TAG No.

| 機能 FUNCTION | レンジ RANGE | 標準器入力 INPUT | 許容差 TOLERANCE | 結果 RESULT |
|----------------|---------------|--------------------------------|------------------|--------------|
| DCV | 50mV | 50mV | 49.965-50.035 | 良 GOOD |
| | 500mV | 500mV | 499.88-500.12 | 良 GOOD |
| | 2400mV | 2000mV | 1,999.4-2,000.6 | 良 GOOD |
| | 5V | 5V | 4.9983-5.0017 | 良 GOOD |
| | 50V | 50V | 49.983-50.017 | 良 GOOD |
| | 500V | 500V | 499.83-500.17 | 良 GOOD |
| | 1000V | 1000V | 999.5-1,000.5 | 良 GOOD |
| ACV | 50mV | 50mV 50Hz | 49.760-50.240 | 良 GOOD |
| | 500mV | 500mV 50Hz | 497.70-502.30 | 良 GOOD |
| | 5V | 5V 50Hz | 4.9770-5.0230 | 良 GOOD |
| | 50V | 50V 50Hz | 49.770-50.230 | 良 GOOD |
| | 500V | 500V 50Hz | 497.70-502.30 | 良 GOOD |
| | 1000V | 1000V 50Hz | 993.0-1,007.0 | 良 GOOD |
| DCA | 500 μ A | 500 μ A | 498.95-501.05 | 良 GOOD |
| | 5000 μ A | 5000 μ A | 4,989.5-5,010.5 | 良 GOOD |
| | 50mA | 50mA | 49.895-50.105 | 良 GOOD |
| | 500mA | 400mA | 399.15-400.85 | 良 GOOD |
| | 5A | 5A | 4.9690-5.0310 | 良 GOOD |
| | 10A | 10A | 9.935-10.065 | 良 GOOD |
| ACA | 500 μ A | 500 μ A 50Hz | 496.05-503.95 | 良 GOOD |
| | 5000 μ A | 5000 μ A 50Hz | 4,960.5-5,039.5 | 良 GOOD |
| | 50mA | 50mA 50Hz | 49.605-50.395 | 良 GOOD |
| | 500mA | 400mA 50Hz | 396.80-403.20 | 良 GOOD |
| | 5A | 5A 50Hz | 4.9480-5.0520 | 良 GOOD |
| | 10A | 10A 50Hz | 9.880-10.120 | 良 GOOD |
| Ω | 500 Ω | 500 Ω | 499.73-500.27 | 良 GOOD |
| | 5k Ω | 5k Ω | 4.9973-5.0027 | 良 GOOD |
| | 50k Ω | 50k Ω | 49.973-50.027 | 良 GOOD |
| | 500k Ω | 500k Ω | 499.73-500.27 | 良 GOOD |
| | 5M Ω | 5M Ω | 4.9748-5.0252 | 良 GOOD |
| | 50M Ω | 50M Ω | 49.498-50.502 | 良 GOOD |
| Cap | 5nF | 5nF | 4.945-5.055 | 良 GOOD |
| | 50nF | 50nF | 49.45-50.55 | 良 GOOD |
| | 500nF | 500nF | 494.5-505.5 | 良 GOOD |
| | 5 μ F | 5 μ F | 4.945-5.055 | 良 GOOD |
| | 50 μ F | 50 μ F | 49.45-50.55 | 良 GOOD |
| | 500 μ F | 500 μ F | 489.5-510.5 | 良 GOOD |
| | 5mF | 5mF | 4.845-5.155 | 良 GOOD |
| | 50mF | 40mF | 38.75-41.25 | 良 GOOD |
| FREQ | 9.999Hz | 9.000Hz | 8.998-9.002 | 良 GOOD |
| | 99.99kHz | 90.00kHz | 89.98-90.02 | 良 GOOD |
| $^{\circ}$ C | - | 800 $^{\circ}$ C+AMBIENT TEMP. | *** | 良 GOOD |

| 項目 ITEM | 結果 RESULT |
|------------------------------|--------------|
| 耐電圧 * DIELECTRIC STRENGTH | 良 GOOD |
| ダイオードテスト DIODE TEST | 良 GOOD |
| 導通チェック CONTINUITY CHECK | 良 GOOD |
| ピークホールド PEAK HOLD | 良 GOOD |
| DCV+ACV DCV+ACV | 良 GOOD |
| DCA+ACA DCA+ACA | 良 GOOD |
| MEAN MEAN | 良 GOOD |
| LP- Ω LP- Ω | 良 GOOD |
| FILTER FILTER | 良 GOOD |
| 外観 APPEARANCE | 良 GOOD |

NOTES

*
入力端子-ケース間 6,880V AC, 5秒間
Between INPUT Terminals and Case 6,880V AC 5 sec.
**
500 Ω ~ 500k Ω , 5nF ~ 50 μ F
レンジは、ゼロ調整後の確度
500 Ω ~ 500k Ω , 5nF ~ 50 μ F
ranges: Accuracy is specified after zero adjustment.

総合結果
RESULT

NOTES ***

(800 $^{\circ}$ C+室内温度) \pm 1% of reading \pm 1.5 $^{\circ}$ C
(800 $^{\circ}$ C+AMBIENT TEMP.) \pm 1% of reading \pm 1.5 $^{\circ}$ C

日付
DATE

室内温度・湿度
AMBIENT TEMP. & HUM.

$^{\circ}$ C & %

検査者
INSPECTED BY

承認者
APPROVED BY

KY